

BIOPSIBLADE

[Contraindications and Prohibitions]

<Applications (Applicable Patients)>

Do not use the product on any patients who are sensitive or allergic to metal.

<Prohibitions>

Re-use prohibited


Re-sterilize prohibited

Do not use the product when the aseptic condition of the products is considered to be lost due to damage/breakage of package.

[Shape and Structure]

This product is a disposable surgical blade which consists of blade and handle.

1. Physical Appearance



2. Material

Blade: Stainless Steel (Chromium contained)

Coating: Chromium, Polytetrafluoroethylene

[Purpose]

This product is a blade for surgery. This product is single use.

[Handling and Use]

Push the blade against the subject tissue, curettage with the appropriate strength and cut the tissue off.

① Before Use

Confirm there is no damage on the sterilized pack, open the package and remove the product.

Confirm there is no damage/breakage of the product and use it without intermission.

② Handling

1. Hold the both ends of the handle by thumb and index finger and put force from both sides to the center, make the blade curved. Keep this form of the blade, cut off the subject tissue.



2. To cut off narrower subject, put additional force from both sides of the blade and make the blade more curved.



[Precautions for Use]

Important Precautions

- (1) Do not use the product if the package is damaged/broken (The aseptic condition cannot be assured even though the product is within the expiry date).
- (2) When removing the product from the package, do this procedure with care of the blade. If the blade gets contact with anything else, do not use the product and replace it with a new one.
- (3) Do not use the product if any abnormality such as damage/breakage is recognized.
- (4) Handle the products with care in order to avoid being injured by the blades (The blade is so sharp that it would injure).
- (5) When handling the product, do not bend it too much (The handle would crack).
- (6) After the usage of the products, be aware of the risk of infection and dispose them properly as medical waste with the perspective of preventing infection.

[Storage and Expiry Date]

1. Storage

- (1) Avoid exposing the product to high temperature, high humidity, direct sunlight and water. Store the product at the ambient temperature.
- (2) Store the product in a clean and dry place without loading anything on the products.
- (3) Do not store the product together with chemical products or in an area where gas occurs.
- (4) Avoid damaging or making pinholes to package materials.

2. Expiry Date

Check the expiry date which is indicated on the individual and external package.

[Manufacturer]

KAI INDUSTRIES CO., LTD.

1110 Oyana, Seki-shi, Gifu 501-3992, Japan