SCALPEL

Do not reuse

[Contraindications and Prohibitions]

< Applications (Applicable Patients) >

Do not use the product on any patients who are sensitive or allergic to metal.

<Prohibitions>

Re-use prohibited

Re-sterilize prohibited

Do not use the product when the aseptic condition of the products is considered to be lost due to damage/breakage of package.

[Shape and Structure]

This product is a disposable surgical scalpel which consists of blade and handle.

1. Physical Appearance

Blade



2. Main Blade Shapes





3. Material

Blade: Stainless Steel (Chromium contained)

[Purpose]

This product is a scalpel for surgery. This product is single use.

[Handling and Use]

To use the scalpel, follow the established surgical method.

[Precautions for Use]

Important Precautions

- (1) Do not use the product if the package is damaged/broken (The aseptic condition cannot be assured even though the product is within their expiry date).
- (2) When opening the package and removing the cap, do this procedure with care of the blade. If the blade gets contact with anything else, do not use the product and replace it with a new one.
- (3) Do not use the product if any abnormality such as damage/breakage is recognized.
- (4) Handle the products with care in order to avoid being injured by the blades (The blade is so sharp that it would injure).
- (5) After the usage of the products, be aware of the risk of infections and dispose them properly as medical waste with the perspective of preventing infection.

[Storage and Expiration Date]

1. Storage

- (1) Avoid exposing the product to high temperature, high humidity, direct sunlight and water. Store the product at the ambient temperature.
- (2) Store the product in a clean and dry place without loading anything on the products.
- (3) Do not store the product together with chemical products or in an area where gas occurs.
- (4) Avoid damaging or making pinholes to package materials.

2. Expiration Date

Check the expiry date which is indicated on the individual and external package.

[Manufacturer]

KAI INDUSTRIES CO., LTD. 1110 Oyana, Seki-shi, Gifu 501-3992, Japan